



广微测
Gmicro Testing



20181900883




GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report No.

2024FM13955R02Da





REPORT FOR ANALYSIS

C5QYVE25

Entrustment Test

	A	1405-07
Address	1405-7, building a, Huaqiang North Street building, Futian District, Shenzhen City	





GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No. 2024FM13955R02Da

1. 1cm 30min
Sample pretreatment: The bacterial carrier was prepared with glass. During the test, the bacterial carrier was placed 1 cm away from the UV lamp and acted for 30 minutes.

2. Test results:

Test germ	Test group	Average cfu of positive controls cfu/	Average cfu of testing groups cfu/	Killing rate %	Sterilization logarithm
CMCC(B) 50115	1	6.6×10^5	5	99.99	5.12
	2	6.9×10^5	5	99.99	5.14
	3	6.8×10^5	5	99.99	5.13

sample picture

Blank below



广 微 测
Gmicro Testing

